

OSP
Osteosynthesis plates (Slotplates)

ENGLISH

THESE INSTRUCTIONS FOR USE MUST BE
READ CAREFULLY PRIOR TO CLINICAL USE

ENGLISH

Intended use

- Surgi-Tec osteosynthesis plates are used in orthognathic and orthofacial surgery. OSPs consists of a plate fixed with monocortical screws. Different models accommodate specific anatomical conditions
- The Osteosynthesis plates are intended for long term use, are invasive , are not active and single-use
- Osteosynthesis plates are to be implanted on the buccal bony surface using dedicated surgical instrumentation and suitable orthosynthesis screw

Intended users

- Suitably trained and qualified oral and maxillo-facial surgeons

Intended patient population

- Patients needing fixation of bone segments after zygoma, maxillary, mandibular and chin osteotomies or trauma.

Intended environment

- The intended environment for the use of osteosynthesis plates is Hospitals and Clinics.

Performance characteristics

- Osteosynthesis plates provide stable anchorage during orthognathic surgery

Material information

Osteosynthesis plates are made of titanium grade 1

Composition of Titanium grade 1

Carbon	Iron	Oxygen	Nitrogen	Hydrogen	Titanium
Max 0,08 %	Max 0,20%	Max 0,18%	Max 0,03%	Max 0,015%	Balance

Indications

- Fixation of the segments after zygoma, maxillary, mandibular and chin osteotomies.
- Fixation of the segments after trauma

Contra indications

- Patients with pre-existing or suspected infection at or near the implantation site
- Patients with known allergies and/or hypersensitivity to implant materials
- Limited blood supply and inferior or insufficient mandibular or maxillar bone quality to hold the fixation screws
- Patients with a history of immune deficiency, steroid therapy, problems with blood clotting, uncontrolled endocrinological disease, rheumatic disease, bone disease, diabetic problems or any other systemic or acute disease
- The OSP must not be used if the patient receives radiotherapy of the head

Possible adverse effects

- Loosening of the implant from loosening of screws or screw failure.
- Hypersensitivity to metal or allergic reactions.
- Difficult removal due to bone overgrowth, premature plate removal.
- Mechanical failure (breakage)
- Morbidities related to orthognathic or orthofacial surgery:
 - Soft tissue irritation and/or nerve damage through surgical trauma
 - Early or late infection, both superficial and deep.
 - Elevated fibrotic tissue reaction around the surgical area
 - Hypo- or dyesthesia / hyposensibility
 - Postsurgical swelling.
 - Wound dehiscence
 - Pseudo-arthrosis
 - Postoperative haemorrhage
 - Frequent maxillary sinusitis,

OSP
Osteosynthesis plates (Slotplates)

ENGLISH

THESE INSTRUCTIONS FOR USE MUST BE
READ CAREFULLY PRIOR TO CLINICAL USE

ENGLISH

- Condylar resorption,
- TMJ problems,
- Instable maxilla,
- Tight intraoral scar,
- Phonetic problems,
- Obstruction of lacrimal canal after maxillary osteotomy
- Nasal bleeding
- Prolonged fatigue or pain postoperatively
- False aneurysms
- Arteriovenous fistulas.

Warnings and precautions

- Read all available documents before first use.
- Check the packaging for integrity, do not use if package is damaged.
Never use products that have been damaged by transport or improper handling.
- Products labelled as 'non sterile' are already cleaned and delivered in non-sterile see-through pouches. They only have to be sterilized in their original packaging before use.
- Products labelled as 'sterile' are gamma irradiated and delivered in sterile see-through pouches. They are ready to be used unless the original packaging has been damaged. If the package is damaged please notify the manufacturer immediately. Do not use the medical device after the use by date indicated on the Labels. Store the sterile device in a dry and dust-free environment
- Osteosynthesis plates may only be used by medical personnel who hold the correct professional qualifications and are familiarized with the procedure.
- Follow appropriate treatment protocols for patients who are on anticoagulant, antiplatelet or aspirin therapy
- Treatment group: adults and children, but a surgeon must always rely on his/her clinical judgement when deciding whether to use a particular product when treating a particular patient
- It is necessary to explain to patients the risks of plate removal and the importance of long-term follow-up
- Inform the patient regarding the possible adverse effects.
- Surgi-Tec does not recommend a specific surgical procedure for a specific patient.
The operating surgeon is responsible for choosing the appropriate Osteosynthesis plate for each specific case.
- Surgi-Tec advises to use only suitable Surgi-Tec screws and related Surgi-Tec instruments.
Insertion of 70-200 series of screws at sharp angles could cause soft tissue irritation or pain.
Repeatedly bending the plate in opposite directions may cause the plate to break during or after surgery.
- Osteosynthesis plates are intended for single use and may not be reused
Re-use may compromise the structural integrity of the device and may create a risk of contamination due to the transmission of infectious material between patients. This could result in injury of the patient or user
- The osteosynthesis plates can be removed.
Implants that were used in a patient and removed, have to be disposed as medical waste in a dedicated container, in accordance with all local guidelines and/or your institution's safety program

Cleaning and Disinfection of Products labelled as 'non-sterile'

- All implants in the Surgi-Tec Osteosynthesis plates portfolio that are delivered '**non-sterile**' must be cleaned, disinfected and sterilized before use. This also applies to the first use after delivery. All packaging must be removed before preparation.
- Thorough cleaning and disinfection are essential for effective sterilization.
- All implant components are intended for one single application in a single patient.
- It is your responsibility to ensure that the implants are completely sterile when used, to use device- and product-specific procedures for cleaning/disinfection and sterilization that are sufficiently validated, to regularly service and inspect the employed devices (disinfector, sterilizer), and to ensure that the validated and/or manufacturer's recommended parameters are maintained for each cycle.
- The statutory regulations applicable in your country and the hospital's hygiene requirements must also be observed. This applies in particular to the various instructions for effectively deactivating prions.
- Surgi-Tec has used "Neodisher MediClean forte" for the validation process of the automated cleaning and disinfection and has followed the instructions of the manufacturer (instruction Dr. Weigert). The validation was carried out according to table below.
- For the remainder of this document, please use the following definitions regarding water temperature:
Cold water: T < 40°C
Warm water: T > 40°C
- When selecting the disinfectant, make sure that the cleaning process includes the following phases in accordance with EN ISO 15883:

OSP
Osteosynthesis plates (Slotplates)

ENGLISH	THESE INSTRUCTIONS FOR USE MUST BE READ CAREFULLY PRIOR TO CLINICAL USE	ENGLISH
----------------	--	----------------

- The following precleaning steps will be performed: Rinsing the test items under cold running tap water for 5 minutes; NOTE: the test items should NOT be disassembled and screwable parts should NOT be moved

Phase	Temperature	Duration	Action
Pre-rinsing	Not applicable	Soaking time > 120 sec / 2 min	Cold tap water (temperature <40°C)
Cleaning cycle	Cleaning temperature - 45°C	Soaking time > 300 sec / 5 min	Warm tap water (temperature >40°C) Recommended Detergent - Neodisher Mediclean Forte Concentration - 0.6% (v/v)
Cleaning cycle	Cleaning temperature - 55°C	Soaking time > 300 sec / 5 min	Warm tap water (temperature >40°C) Recommended Detergent - Neodisher Mediclean Forte Concentration - 0.6% (v/v)
Post-rinsing 1	Rinsing temperature >40°C	Rinsing time > 60 sec / 1 min	Warm tap water (temperature >40°C)
Post-rinsing 2	Rinsing temperature <40°C	Rinsing time > 60 sec / 1 min	Cold tap water (temperature <40°C)
Thermal disinfection (A ₀ value > 600)* (A ₀ value > 3000)	Disinfection temperature ≥ 90°C (194°F)	Soaking time > 60 sec / 1 min (A ₀ value > 600) > 300 sec / 5 min (A ₀ value > 3000)	With demineralized (DI water) and/or purified water; do not add additional detergent
Drying	Drying temperature > 110°C	Drying time > 1500 sec / 25 min*	Drying process
(*) Corresponds to worst case item validated			

- The information provided is based on the use of “Neodisher MediClean forte” by Dr. Weigert; validation was performed with a concentration of 0.60 % v/v at 55°C; if a different detergent is used, exposure times, concentrations and temperatures may vary; the relevant manufacturer’s instructions must be observed.
- The products must be completely dried directly afterwards. It is recommended to dry the products using medical compressed air; this is especially gentle and effective. Otherwise lint-free disposable wipes (e.g. Perform classic from Schülke & Mayr) can be used. If applicable, the products have to be stored in a clean environment until they are completely dry.

Inspection

- In general sufficient cleanliness is the basic requirement for a successful sterilization. Before the products are packaged for sterilization they have to be inspected visually. (Recommendation: use working place light fixtures ideally with magnifiers).
- Check the Osteosynthesis plates and accessories after cleaning and disinfection for damages and contamination.

Packaging

- Surgi-Tec recommends Single sterilization wrapping (single or double wrapping) and/or other sterilization containers can be used.

Sterilization of Products labelled as ‘non sterile’

- Surgi-Tec TPD osteosynthesis plates are made of Titanium Grade 1
- For the sterilization process the instructions of the appropriate sterilizers have to be followed.
- All NON-STERILE products can be sterilized in an autoclave. The autoclaves must be in accordance with EN285 respectively EN13060 regarding validation, servicing, maintenance and controlling.
- Steam sterilization should be performed after the recommended cleaning, disinfection, inspection and packaging.
- Surgi-Tec recommends that sterilization is performed in accordance with following EN ISO 17665 validated process parameters

OSP
Osteosynthesis plates (Slotplates)

ENGLISH	THESE INSTRUCTIONS FOR USE MUST BE READ CAREFULLY PRIOR TO CLINICAL USE	ENGLISH
----------------	--	----------------

- For both initial and subsequent sterilization, the following parameters were validated by Surgi-Tec in accordance with the requirements of the current sterilization standards, EN ISO 17665 and ANSI/AAMI ST79.

- **Cycle: Pre-Vacuum (Dynamic air removal)**

Procedure	Fractionated and dynamic pre-vacuum process
Exposure time	≥ 5 minutes <i>(Minimum 5 minutes at 134°C)</i>
Temperature	Temperature ≥ 134°C
Drying time	≥ 20 minutes - 30 minutes (in autoclave chamber)

- Surgi-Tec recommends that sterilization is performed in accordance with the above validated processes. If the user utilizes other processes (e.g. flash sterilization), these must be validated by the user. The ultimate responsibility for validation of sterilization techniques and equipment lies with the user.
- The responsibility for the maintenance and qualification of the sterilization equipment used and the validation of user specific sterilization process lies with the user.
- After sterilization, the products must be stored in a dry and dust-free environment. The maximum storage time is dependent on different factors such as the sterile barrier employed, storage manner, environmental conditions and handling.
- The user should define a maximal storage time for sterile products until use. Within this defined time the products have to be used or reprocessed again.
- Do not use hot-air sterilization, radiation sterilization, formaldehyde sterilization, ethylene oxide sterilization or substitute procedures for sterilizing thermolabile products such as plasma or peroxide sterilization for Surgi-Tec products.

These IFU and additional information can be found on the internet at www.surgi-tec.com/instructions.

The summary of safety and performance can be found in Eudamed public website:

<https://ec.europa.eu/tools/eudamed>














**OSP
Osteosynthesis plates (Slotplates)**

ENGLISH

THESE INSTRUCTIONS FOR USE MUST BE
READ CAREFULLY PRIOR TO CLINICAL USE

ENGLISH

Explanation of symbols

	Please observe instructions for use
	Caution
	Do Not Re-use
	Reference number
	Lot number
	Manufacturer
	Non sterile product
	Do not use if package is damaged
	Medical device
	Unique device identifier
	Sterilized using irradiation
	Use-by date
	Double sterile barrier system

OSP
Osteosynthesis plates (Slotplates)

ENGLISH

THESE INSTRUCTIONS FOR USE MUST BE
READ CAREFULLY PRIOR TO CLINICAL USE

ENGLISH



1639

Medical device Class II.b

BASIC UDI-DI : 540700775OSPH7

MANUFACTURED BY
"SURGI-TEC"
Poortakkerstraat 43
9051 SINT-DENIJS-WESTREM – BELGIUM
www.surgi-tec.com